

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU
										164.0	226.0	20.5	0.7	0.0
400301351168	03/23/2019	04/30/2019	Scheduled Service	MM	In Progress	PART1&2 - DCS&SUPPORTAR...	-	0	N	0.0	0.0	0.0	0.0	0.0
400109414974	04/23/2019	04/24/2019	Scheduled Service	MS	Completed	SRSC-EPS GUARDIAN, activat...	N	1	Y	3.0	2.0	0.0	0.0	0.0
400109403427	04/16/2019	04/17/2019	Partially Down	MS	Completed	Need assist with log in after log off an	Y	0	N	0.0	0.0	0.5	0.0	0.0
400301315206	09/22/2018	04/09/2019	Scheduled Service	MM	Completed	PART 4 10Y Annual	-	1	N	2.5	5.5	0.0	0.0	0.0
400109387347	04/06/2019	04/07/2019	Partially Down	MS	Completed	Turned off system the wrong way, passwor	Y	0	N	0.0	0.0	0.9	0.0	0.0
400780218719	04/26/2018	03/26/2019	Scheduled Service	MA	Completed	UI-AX028/17/P APPS SW MAINTENANCE...	-	0	N	0.0	0.0	0.0	0.0	0.0
400501237983	04/25/2018	03/20/2019	Scheduled Service	MU	Completed	UI-AX028/17/P APPS SW MAINTENANCE...	-	0	N	0.0	0.0	0.0	0.0	0.0
400109351685	03/18/2019	03/19/2019	Partially Down	MS	Completed	requests meduser name and password	N	1	Y	2.0	1.5	0.4	0.0	0.0
400501254454	06/21/2018	02/15/2019	Scheduled Service	MU	Completed	UI-AX003/16/S HW-UPDATE ALGAE	-	0	N	0.0	0.0	0.0	0.0	0.0
400780256082	01/30/2019	02/14/2019	Scheduled Service	MA	Completed	Self-study program	-	0	N	0.0	0.0	0.0	0.0	0.0
400780249070	01/30/2019	02/07/2019	Scheduled Service	MA	Completed	e.learning subscription	-	0	N	0.0	0.0	0.0	0.0	0.0

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU
										164.0	226.0	20.5	0.7	0.0
400780249005	01/30/2019	01/30/2019	Scheduled Service	MA	In Progress	e.learning subscription	-	0	N	0.0	0.0	0.0	0.0	0.0
400780254887	01/30/2019	01/30/2019	Scheduled Service	MA	In Progress	Self-study program	-	0	N	0.0	0.0	0.0	0.0	0.0
400109255202	01/22/2019	01/25/2019	Partially Down	MS	Completed	Floor stand not locking in place when mo	N	1	Y	1.0	3.0	0.3	0.0	0.0
400109209497	12/22/2018	12/25/2018	Partially Down	MS	Completed	emergency stop button was activated...	N	1	Y	3.0	1.5	0.5	0.0	0.0
400108988638	08/16/2018	12/15/2018	Scheduled Service	MS	Completed	SRSC-EPS GUARDIAN, system...	N	1	Y	0.0	0.5	0.0	0.0	0.0
400501237902	04/24/2018	12/07/2018	Scheduled Service	MU	Completed	UI-AX025/17/P SW-UPDATE VC14J MA...	-	1	N	2.5	12.0	0.0	0.0	0.0
400301214455	07/31/2017	10/03/2018	Scheduled Service	MM	Completed	PART3+DCS&SUPPO RTARM(P2) 8Y 9 M...	-	1	N	3.0	6.0	0.0	0.0	0.0
400108878959	06/14/2018	09/25/2018	Partially Down	MS	Completed	Per RSM Sean Casey - image quality and d	Y	0	N	0.0	0.0	1.0	0.0	0.0
400109048626	09/19/2018	09/22/2018	Partially Down	MS	Completed	CSE 42584 reported joystick is broken	N	1	Y	0.0	0.5	0.0	0.0	0.0
400109042406	09/17/2018	09/22/2018	Partially Down	MS	Completed	System is loosing time.	N	2	N	2.5	2.0	1.5	0.0	0.0
400301297514	06/23/2018	09/20/2018	Scheduled Service	MM	Completed	PART3+DCS&SUPPO RTARM(P2) 9Y 9 M...	-	1	N	1.5	5.5	0.0	0.0	0.0
400109042423	09/17/2018	09/18/2018	Partially Down	MS	Completed	Time display on system is not accurate	Y	0	N	0.0	0.0	0.0	0.1	0.0
400109019099	09/04/2018	09/05/2018	Room Down	MS	Completed	The battery operated peddle did not work	N	2	N	6.0	2.5	0.0	0.0	0.0
400780229947	08/31/2018	09/03/2018	Scheduled Service	MA	Task Outstanding	Self-study program	-	0	N	0.0	0.0	0.0	0.0	0.0
400780230213	08/31/2018	08/31/2018	Scheduled Service	MA	In Progress	e.learning subscription	-	0	N	0.0	0.0	0.0	0.0	0.0
400108871195	06/11/2018	07/08/2018	Partially Down	MS	Completed	Monitor is out, no power going to it	N	2	N	5.0	3.5	0.0	0.0	0.0

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU	
											164.0	226.0	20.5	0.7	0.0
400301274389	03/30/2018	06/29/2018	Scheduled Service	MM	Completed	PART1&2 - DCS&SUPPORTAR...	-	1	N	1.5	0.0	0.0	0.0	0.0	
400108854056	05/31/2018	06/29/2018	Scheduled Service	MS	Completed	SRSC-EPS GUARDIAN, re-regi...	N	1	Y	1.0	2.0	0.0	0.0	0.0	
400108863740	06/06/2018	06/08/2018	Partially Down	MS	Completed	Edit exam sets	Y	0	N	0.0	0.0	0.0	0.4	0.0	
400108861033	06/05/2018	06/07/2018	Partially Down	MS	Completed	Hemo dynamics monitor not working	N	1	Y	3.0	1.5	0.8	0.0	0.0	
400108840576	05/23/2018	06/02/2018	Scheduled Service	MS	Completed	SRSC-EPS GUARDIAN, re-regi...	Y	0	N	0.0	0.0	0.0	0.0	0.0	
400501172836	08/24/2017	05/09/2018	Scheduled Service	MU	Completed	UI-AX001/17/S HW-WIRELESS FOOT...	-	1	N	0.0	1.0	0.0	0.0	0.0	
400501185192	09/19/2017	05/09/2018	Scheduled Service	MU	Completed	UI-AX009/17/S HW-UPDATE WIRELES...	-	1	N	0.0	1.0	0.0	0.0	0.0	
400501127013	06/02/2017	05/09/2018	Scheduled Service	MU	Completed	UI-AX043/16/S HW-UPDATE TABLE LE...	-	2	N	0.0	0.5	0.0	0.0	0.0	
400501212358	12/11/2017	05/09/2018	Scheduled Service	MU	Completed	UI-AX064/17/S INSPECTION FOR...	-	1	N	1.5	0.5	0.0	0.0	0.0	
400108743258	03/30/2018	04/04/2018	Partially Down	MS	Completed	Reduces stand speed, cannot move stand i	Y	0	N	0.0	0.0	0.2	0.0	0.0	
400108700821	03/07/2018	03/24/2018	Scheduled Service	MS	Completed	Physicist reports system dose is not cor	N	2	N	3.5	3.0	0.3	0.0	0.0	
400108613123	01/18/2018	01/26/2018	Partially Down	MS	Completed	Had to shut down system to replace b...	N	3	N	9.5	6.0	0.6	0.0	0.0	
400108535808	12/04/2017	01/20/2018	Partially Down	MS	Completed	scan abort. Room available now	N	3	N	3.0	3.5	0.6	0.0	0.0	
400301226906	09/22/2017	01/04/2018	Scheduled Service	MM	Completed	PART 4 9Y Annual	-	1	N	3.5	5.5	0.0	0.0	0.0	
400501209300	12/08/2017	12/15/2017	Scheduled Service	MU	Completed	UI-AX065/17/S CSI for AX064/17/S Inspect	-	0	N	0.0	0.0	0.0	0.0	0.0	
400501005044	02/17/2016	12/14/2017	Scheduled Service	MU	Completed	UI-AX004/15/S HW-UPDATE LAIRD FD...	-	1	N	4.0	3.5	0.0	0.0	0.0	

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU
										164.0	226.0	20.5	0.7	0.0
400108535319	12/04/2017	12/06/2017	Partially Down	MS	Completed	Tube too hot	N	1	Y	3.0	7.0	0.2	0.0	0.0
400108479815	11/01/2017	11/11/2017	Partially Down	MS	Completed	Limited stand/table movement	N	1	Y	1.5	1.0	0.9	0.0	0.0
400108482064	11/02/2017	11/03/2017	Partially Down	MS	Completed	Foot paddle is OUT and will not charge.	N	1	Y	2.5	2.5	0.6	0.0	0.0
400108466204	10/24/2017	10/30/2017	Partially Down	MS	Completed	Unable to transfer images to PACS	N	1	Y	2.5	2.0	0.6	0.0	0.0
400501187397	09/19/2017	09/29/2017	Scheduled Service	MU	Completed	UI-AX013/17/S CSI FOR AX009/17/S W...	-	0	N	0.0	0.0	0.0	0.0	0.0
400108381605	09/08/2017	09/16/2017	Partially Down	MS	Completed	Stand is moving slow	N	2	N	3.0	5.5	0.4	0.0	0.0
400501174932	08/24/2017	09/02/2017	Scheduled Service	MU	Completed	UI-AX002/17/S CSAN for AX001/17/S	-	0	N	0.0	0.0	0.0	0.0	0.0
400501132869	06/08/2017	08/26/2017	Scheduled Service	MU	Completed	UI-AX038/17/S WinXP SMB Vulnerability	-	1	N	1.0	1.0	0.0	0.0	0.0
400108306870	07/28/2017	08/05/2017	Partially Down	MS	Completed	Doctor thinks dose may be reporting high	N	1	Y	1.0	1.0	0.3	0.0	0.0
400301185881	03/23/2017	08/05/2017	Scheduled Service	MM	Completed	PART 2 8Y 6 Months	-	1	N	1.5	6.5	0.0	0.0	0.0
400301206047	06/23/2017	08/05/2017	Scheduled Service	MM	Completed	PART 3 8Y 9 Months	-	0	N	0.0	0.0	0.0	0.0	0.0
400108274275	07/12/2017	07/15/2017	Partially Down	MS	Completed	Control room screen is flickering, unabl	N	2	N	6.0	3.0	0.5	0.0	0.0
400108268618	07/09/2017	07/15/2017	Partially Down	MS	Completed	would not take any xray, said shutdown r	N	1	Y	7.0	1.0	0.4	0.0	0.0
400501136326	06/08/2017	06/17/2017	Scheduled Service	MU	Completed	UI-AX037/17/S CSI SMB Vulnerability	-	0	N	0.0	0.0	0.0	0.0	0.0
400501129733	06/02/2017	06/09/2017	Scheduled Service	MU	Completed	UI-AX065/16/S CSI AX043/16/S TABLE...	-	0	N	0.0	0.0	0.0	0.0	0.0
400108210320	06/06/2017	06/07/2017	Partially Down	MS	Completed	Wireless footswitch is not charging. Wor	N	1	Y	3.0	1.0	0.1	0.0	0.0

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU
										164.0	226.0	20.5	0.7	0.0
400108151629	05/04/2017	05/06/2017	Partially Down	MS	Completed	Footswitch only working if plugged in...	N	1	Y	2.5	1.5	0.2	0.0	0.0
400501096506	02/20/2017	04/08/2017	Scheduled Service	MU	Completed	UI-AX063/16/S CSAN FOR WIRELESS F...	-	0	N	0.0	0.0	0.0	0.0	0.0
400301165058	12/22/2016	03/28/2017	Scheduled Service	MM	Completed	PART 1 8Y 3 Months	-	1	N	1.5	5.0	0.0	0.0	0.0
400501043740	07/13/2016	03/25/2017	Scheduled Service	MU	Completed	UI-AX006/15/S HW-ARTIS TUBE COOL...	-	1	N	1.0	1.0	0.0	0.0	0.0
400108038397	03/02/2017	03/04/2017	Partially Down	MS	Completed	Fluoro Check - Did not have to move C-Ar	N	1	Y	3.5	2.5	0.4	0.0	0.0
400501090432	02/07/2017	02/25/2017	Scheduled Service	MU	Completed	UI-AX058/16/S CSI /ADDENDUM F...	-	0	N	0.0	0.0	0.0	0.0	0.0
400107937557	01/06/2017	01/28/2017	Partially Down	MS	Completed	Startup failed, please call service	N	7	N	19.0	34.5	2.6	0.0	0.0
400780163797	11/29/2016	01/18/2017	Scheduled Service	MA	In Progress	Self-study program	-	0	N	0.0	0.0	0.0	0.0	0.0
400107922924	12/28/2016	12/31/2016	Room Down	MS	Completed	System startup failed	Y	0	N	0.0	0.0	0.3	0.0	0.0
400301145135	09/23/2016	12/24/2016	Scheduled Service	MM	Completed	PART 4 8Y Annual	-	1	N	1.0	5.5	0.0	0.0	0.0
400501068224	10/06/2016	12/24/2016	Scheduled Service	MU	Completed	UI-AX028/16/P SW-Update VC14J Patc...	-	2	N	1.0	1.0	0.0	0.0	0.0
400107780392	10/11/2016	10/15/2016	Partially Down	MS	Completed	Needs assistance in merging images.	Y	0	N	0.0	0.0	0.0	0.3	0.0
400107778115	10/10/2016	10/15/2016	Partially Down	MS	Completed	Xray intermittent.	N	2	N	4.0	4.5	0.7	0.0	0.0
400301124167	06/23/2016	09/24/2016	Scheduled Service	MM	Completed	ART 3 7Y 9 Months	-	1	N	1.0	6.5	0.0	0.0	0.0
400501034638	06/17/2016	09/24/2016	Scheduled Service	MU	Completed	UI-AX071/15/P INSPECTION OF D...	-	2	N	1.5	1.0	0.0	0.0	0.0
400107731257	09/14/2016	09/17/2016	Partially Down	MS	Completed	HIS/RIS connection error	Y	0	N	0.0	0.0	0.4	0.0	0.0

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU
										164.0	226.0	20.5	0.7	0.0
400107641988	07/26/2016	08/20/2016	Scheduled Service	MS	Completed	SRSC-EPS GUARDIAN, not RU...	N	1	Y	1.5	2.0	0.0	0.0	0.0
400501046931	07/13/2016	08/20/2016	Scheduled Service	MU	Completed	UI-AX007/15/S CSI FOR AX006/15/S	-	0	N	0.0	0.0	0.0	0.0	0.0
400301102901	03/23/2016	07/02/2016	Scheduled Service	MM	Completed	PART 2 7Y 6 Months	-	2	N	1.0	5.0	0.0	0.0	0.0
400107576648	06/20/2016	06/25/2016	Partially Down	MS	Completed	Charge station damaged by users	N	1	Y	1.5	0.5	0.2	0.0	0.0
400107582623	06/22/2016	06/25/2016	Scheduled Service	MS	Completed	SRSC-EPS reinstall system managemen...	N	1	Y	1.5	2.0	1.0	0.0	0.0
400780106390	03/11/2015	06/24/2016	Scheduled Service	MA	Task Outstanding	e.learning subscription	-	0	N	0.0	0.0	0.0	0.0	0.0
400780143544	04/29/2016	06/24/2016	Scheduled Service	MA	In Progress	Self-Study Program	-	0	N	0.0	0.0	0.0	0.0	0.0
400780106379	03/11/2015	06/24/2016	Scheduled Service	MA	Task Outstanding	Self-study program	-	0	N	0.0	0.0	0.0	0.0	0.0
400107545107	06/02/2016	06/18/2016	Partially Down	MS	Completed	Tube hot message.	N	3	N	8.0	20.5	0.6	0.0	0.0
400107528704	05/24/2016	05/28/2016	Partially Down	MS	Completed	Slow Carm movements	N	2	N	6.0	3.5	0.5	0.0	0.0
400107396958	03/11/2016	03/19/2016	Partially Down	MS	Completed	AX_CAN_313 - COLL_N_BASE_V3:...	N	1	Y	1.0	3.0	0.0	0.0	0.0
400301081489	12/22/2015	03/19/2016	Scheduled Service	MM	Completed	PART 1. 7Y 3 Months	-	1	N	1.0	5.0	0.0	0.0	0.0
400107251850	12/22/2015	03/19/2016	Partially Down	MS	Completed	SRSC-EPS problem with UM reports, ple...	N	1	Y	0.0	1.5	1.1	0.0	0.0
400107373954	02/29/2016	03/19/2016	Scheduled Service	MS	Completed	top of the casing is cracked in the corn	N	3	N	1.0	1.0	0.0	0.0	0.0
400501006801	02/17/2016	02/27/2016	Scheduled Service	MU	Completed	UI-AX005/15/S CSI FOR AX004/15/S F...	-	0	N	0.0	0.0	0.0	0.0	0.0
400107347645	02/15/2016	02/20/2016	Partially Down	MS	Completed	cse 42584	N	1	Y	2.5	1.5	0.0	0.0	0.0

SERVICE HISTORY REPORT

Notification	Created	Last Changed	Effect Code	Type	Status	Description	Remote	OSRs	FVFR	TRV	WRK	TSE	APP	EDU
										164.0	226.0	20.5	0.7	0.0
400107228704	12/09/2015	01/30/2016	Scheduled Service	MS	Completed	cse 42631 customer request installatio	N	2	N	0.0	0.5	0.0	0.0	0.0
400700172669	01/08/2016	01/30/2016	Scheduled Service	MI	Completed	OPT 30191003 C-Room DVI	-	1	N	3.5	3.5	0.0	0.0	0.0
400107224537	12/07/2015	01/16/2016	Partially Down	MS	Completed	Guardian Message: BaseRotA- Stand/ta...	N	2	N	2.5	3.5	0.9	0.0	0.0
400301059278	09/22/2015	10/31/2015	Scheduled Service	MM	Completed	PART 4 7Y Annual	-	2	N	2.5	4.5	0.0	0.0	0.0

SERVICE HISTORY REPORT

Doc #	DocType	Notification	Post Date	Return Post Date	Material #	Description	Movement Type	Debit/Credit	PO Qty
	N/A	4005012379 02	12/06/2018	-	10691716	Docu-EN02- Update VC...	GI for Update	Debit	1
	N/A	4005012379 02	12/06/2018	-	11327659	SW UPDATE VC21 MO...	GI for Update	Debit	1
	N/A	4001090486 26	09/21/2018	-	10280923	ECC-J interconne...	GI for Contract	Debit	1
	N/A	4001090486 26	09/21/2018	-	10280814	ECC operation...	GI for Contract	Debit	1
	N/A	4001090424 06	09/19/2018	-	3082146	LITHIUM BATTERY...	GI for Contract	Debit	2
	N/A	4003012975 14	09/19/2018	-	10864418	SIGNAL LAMP 28V...	GI for Contract	Debit	1
	N/A	4003012975 14	09/19/2018	-	11192199	spare part air filter Angio	GI for Contract	Debit	1
	N/A	4001090190 99	09/04/2018	-	4787797	Wireless XRay FSW...	GI for Contract	Debit	1
	N/A	4001088711 95	07/07/2018	-	10293047	MONITOR BW 19" DS...	GI for Contract	Debit	1
	N/A	4001088610 33	06/06/2018	-	8683984	MONITOR COLOR 19...	GI for Contract	Debit	1
	N/A	4005012123 58	05/08/2018	-	11344339	Bolt Control Set	GI for Update	Debit	1

SERVICE HISTORY REPORT

Doc #	DocType	Notification	Post Date	Return Post Date	Material #	Description	Movement Type	Debit/Credit	PO Qty
	N/A	4005011270 13	05/08/2018	-	10280753	HW update Artis table	GI for Update	Debit	1
	N/A	4005011851 92	05/08/2018	-	10892517	HW Update circuit boar...	GI for Update	Debit	1
	N/A	4005011728 36	05/08/2018	-	10892518	HW update seal Wirele...	GI for Update	Debit	1
	N/A	4001086131 23	01/25/2018	-	10520998	ASM VDX board in V...	GI for Contract	Debit	1
	N/A	4001086131 23	01/19/2018	-	10143632	Display labels Artis	GI for Contract	Debit	1
	N/A	4005010050 44	12/13/2017	-	10143910	HW update FD Kluever...	GI for Update	Debit	1
	N/A	4001085358 08	12/07/2017	-	7395143	Set of carbon brushes	GI for Contract	Debit	1
	N/A	4001085358 08	12/07/2017	-	7395168	Set of carbon brushes	GI for Contract	Debit	1
	N/A	4001085358 08	12/07/2017	-	7395176	Set of carbon brushes	GI for Contract	Debit	1
	N/A	4001085353 19	12/05/2017	-	5764555	Kluever cooling unit	GI for Contract	Debit	1
	N/A	4001084798 15	11/08/2017	-	7150998	Potentiomete r 1K 1.5W...	GI Warr Parts- Serv	Debit	3
	N/A	4001084820 64	11/03/2017	-	4787797	Wireless XRay FSW...	GI for Contract	Debit	1
	N/A	4003012144 55	09/26/2017	-	10864418	SIGNAL LAMP 28V...	GI for Contract	Debit	1
	N/A	4001083816 05	09/13/2017	-	7150998	Potentiomete r 1K 1.5W...	GI for Contract	Debit	2
	N/A	4003011858 81	08/01/2017	-	10093773	Disc 4-pole (Claymount)	GI for Contract	Debit	2
	N/A	4003011858 81	08/01/2017	-	10093772	Disc, 3-pole (Claymount)	GI for Contract	Debit	2

SERVICE HISTORY REPORT

Doc #	DocType	Notification	Post Date	Return Post Date	Material #	Description	Movement Type	Debit/Credit	PO Qty
	N/A	4003011858 81	08/01/2017	-	11270288	Filter insert Laird	GI for Contract	Debit	1
	N/A	4003011858 81	08/01/2017	-	10390318	FILTER STRAINER	GI for Contract	Debit	1
	N/A	4003011858 81	08/01/2017	-	1786995	GASKET	GI for Contract	Debit	2
	N/A	4003011858 81	08/01/2017	-	1787035	SILICONOIL AK350	GI for Contract	Debit	2
	N/A	4001082742 75	07/13/2017	-	10293047	MONITOR BW 19" DS...	GI for Contract	Debit	1
	N/A	4001082103 20	06/07/2017	-	4787789	Batteries, wireless X...	GI Warr Parts- Serv	Debit	1
	N/A	4001081516 29	05/05/2017	-	4787789	Batteries, wireless X...	GI for Contract	Debit	1
	N/A	4005010437 40	03/22/2017	-	11270281	External filter Laird / Meg...	GI for Srvc Courtesy	Debit	1
	N/A	4001079375 57	01/26/2017	-	10498048	AXIS IVS si basic PC R...	GI for Contract	Debit	1
	N/A	4001077781 15	10/12/2016	-	10961681	ETHERNET- SWITCH 1...	GI for Contract	Debit	1
	N/A	4003011029 01	06/30/2016	-	10093773	Disc 4-pole (Claymount)	GI for Contract	Debit	2
	N/A	4003011029 01	06/30/2016	-	10093772	Disc, 3-pole (Claymount)	GI for Contract	Debit	2
	N/A	4003011029 01	06/30/2016	-	1786508	DISK 3-pole	GI for Contract	Debit	2
	N/A	4003011029 01	06/30/2016	-	1786516	DISK 4-pole	GI for Contract	Debit	2
	N/A	4003011029 01	06/30/2016	-	10390318	FILTER STRAINER	GI for Contract	Debit	1
	N/A	4003011029 01	06/30/2016	-	1786995	GASKET	GI for Contract	Debit	2

SERVICE HISTORY REPORT

Doc #	DocType	Notification	Post Date	Return Post Date	Material #	Description	Movement Type	Debit/Credit	PO Qty
	N/A	4003011029 01	06/30/2016	-	10548970	PNT.MARKR SIGNALW...	GI for Contract	Debit	2
	N/A	4003011029 01	06/30/2016	-	1787035	SILICONOIL AK350	GI for Contract	Debit	2
	N/A	4003011029 01	06/30/2016	-	10548945	SPRAYPNT SIGNALW...	GI for Contract	Debit	1
	N/A	4001075766 48	06/21/2016	-	4787763	Charg. station, wir...	GI for Contract	Debit	1
	N/A	4001075766 48	06/21/2016	-	8870698	Ladekabel,ex t. (WF)	GI for Contract	Debit	1
	N/A	4001075451 07	06/02/2016	-	11253350	Pump upgrade to...	GI for Contract	Debit	1
	N/A	4001075287 04	05/25/2016	-	6465590	Potentiomete r	GI Warr Parts- Serv	Debit	1
	N/A	4001075287 04	05/25/2016	-	6549948	Potentiomete r, kpl. for s...	GI Warr Parts- Serv	Debit	1
	N/A	4001073739 54	03/18/2016	-	6549096	Receptor cover, left	GI for Contract	Debit	1
	N/A	4003010814 89	03/15/2016	-	4787789	Batteries, wireless X...	GI for Contract	Debit	1
	N/A	4001073739 54	03/15/2016	-	6549104	Receptor cover, right	GI for Contract	Debit	1
	N/A	4001073969 58	03/14/2016	-	5891309	Pre-filter 1-2- 6, Revision...	GI for Contract	Debit	1
	N/A	4001072245 37	12/08/2015	-	6465590	Potentiomete r	GI for Contract	Debit	1
	N/A	4001072245 37	12/08/2015	-	6549948	Potentiomete r, kpl. for s...	GI for Contract	Debit	1